

Certificate of Analysis

ISO 17034

Pesticides Resolution Check Standard

Product Number: CLP-242-1**Page:** 1 of 1**Lot Number:** 0006484858**Lot Issue Date:** 15-Aug-2019**Expiration Date:** 30-Sep-2021

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
trans-chlordane	005103-74-2	RM02726	10.0 ± 0.1 ng/mL
4,4'-DDE	000072-55-9	RM02892	20.0 ± 0.1 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM01256	20.0 ± 0.1 ng/mL
dieldrin	000060-57-1	RM15334	20.1 ± 0.1 ng/mL
endosulfan I	000959-98-8	RM00834	10.0 ± 0.1 ng/mL
endosulfan sulfate	001031-07-8	RM12593	20.0 ± 0.1 ng/mL
endrin ketone	053494-70-5	NT00720	20.0 ± 0.1 ng/mL
methoxychlor	000072-43-5	RM14186	100.1 ± 0.5 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM13844	20.1 ± 0.1 ng/mL

Matrix: hexane**Storage:** Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System



ISO 17025 Cert No.
AT-1937