

Certificate of Analysis

Product Name: ICP-MS Tuning Solution Low Concentration-480mL

Product Number: CICM-699

Lot Issue Date: 19-Nov-2020

Lot Number: 0006571602

Expiration Date: 31-Dec-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO 17034 / ISO 17025 accredited laboratory and compared to calibration standards independently prepared using NIST SRM(s) when available. The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
lithium carbonate (ICP grade) (as lithium)	000554-13-2	RM07634	1.00 ± 0.01 µg/L
yttrium oxide (ICP grade) (as yttrium)	001314-36-9	RM13832	1.00 ± 0.01 µg/L
cerium (III) nitrate (ICP grade) (as cerium)	010294-41-4	RM09055	1.00 ± 0.01 µg/L
thallium (I) nitrate (ICP grade) (as thallium)	010102-45-1	RM10594	1.00 ± 0.01 µg/L
cobalt nitrate hexahydrate (ICP grade) (as cobalt)	010026-22-9	RM10787	1.00 ± 0.01 µg/L
magnesium nitrate hexahydrate (ICP grade) (as magnesium)	013446-18-9	RM13723	1.00 ± 0.01 µg/L

Matrix: water with dilute nitric acid

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.



Certificate of Analysis

Product Number: CICM-699

Lot Number: 0006571602

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-040.1



ISO 17025 Cert
No. AT-1937