

Reference Material Certificate
Product Information Sheet

Product Name: IRSA 5060 Pesticide Mix

Lot Number: 0006683393

Product Number: PST-ITA-02

Lot Issue Date: 17-May-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 30-Jun-2023

Component Name	CERTIFIED VALUES			CAS#	Analyte Lot
	Concentration	Expanded	Uncertainty		
methoxychlor	10.0	±	0.1 mg/L	000072-43-5	RM14186
hexazinone	10.0	±	0.1 mg/L	051235-04-2	RM18243
metazachlor	10.0	±	0.1 mg/L	067129-08-2	RM19530
orbencarb	10.0	±	0.1 mg/L	034622-58-7	RM14240
vinclozolin	10.0	±	0.1 mg/L	050471-44-8	RM09254
sebuthylazine	10.0	±	0.1 mg/L	007286-69-3	RM18729
trifluralin	10.0	±	0.1 mg/L	001582-09-8	RM09256
terbuthylazine-desethyl	10.0	±	0.1 mg/L	030125-63-4	RM19260
atrazine-desisopropyl	10.0	±	0.1 mg/L	001007-28-9	RM14510
atrazine-desethyl	10.0	±	0.1 mg/L	006190-65-4	RM19499
simazine	10.0	±	0.1 mg/L	000122-34-9	RM00622
atrazine	10.0	±	0.1 mg/L	001912-24-9	RM17497
propazine	10.0	±	0.1 mg/L	000139-40-2	RM14562
terbuthylazine	10.0	±	0.1 mg/L	005915-41-3	RM18246
alachlor	10.0	±	0.1 mg/L	015972-60-8	RM14112
prometryn	10.0	±	0.1 mg/L	007287-19-6	RM18571
terbutryn(e)	10.0	±	0.1 mg/L	000886-50-0	RM16747
metolachlor	10.0	±	0.1 mg/L	051218-45-2	RM19646
cyanazine	10.0	±	0.1 mg/L	021725-46-2	RM18558
pendimethalin	10.0	±	0.1 mg/L	040487-42-1	RM18372
molinate	10.0	±	0.1 mg/L	002212-67-1	RM14835
desmetryne	10.0	±	0.1 mg/L	001014-69-3	RM18027
ametryn	10.0	±	0.1 mg/L	000834-12-8	RM18570

Matrix: acetone

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative