Certificate of Analysis

Product Number: ICP-023
Lot Number: 0006684276

Lot Issue Date: 26-May 2022
Expiration Date: 30-Jun 2029

Product Name: Vanadium ICP Standard

Description:
This Reference Material (RM) was gravimetrically prepared in accordance with ISO 17034 and under Agilent's ISO 9001 registered quality system. The neat materials used for this product have been verified by Agilent's ISO 17025 laboratory and under Agilent's ISO 17034 accreditation. The analyte concentrations were verified by Agilent's ISO 17025 accredited laboratory. For each analyte, the true value, with its uncertainty value calculated at the 95% confidence level, is reported below.

Analyte	Starting Material	Lot Number	Purity (%)	Analyte Concentration	Traceability & Method
vanadium	ammonium metavanadate	RM12372	99.999	1001 ± 2 µg/mL	NIST SRM 3165; ICP-OES

Solvent: 5% nitric acid with 1% hydrofluoric acid in low TOC water (< 50 ppb)

Non-Certified Values:

Density: 1.0241 g/mL @ 22.3 ± 0.5°C

Trace Metallic Impurities in Solution Standard in µg/mL

*___	Al	<0.005 ND	*___	Ga	<0.005 ND	n___	Nb		n___	S				
*___	Sb	<0.005 ND	n___	Ge		n___	Os		n___	Ta				
*___	As	<0.005 ND	n___	Au		*___	Pd	<0.005 ND	n___	Te				
*___	Ba	<0.005 ND	n___	Hf		*___	P	<0.005 ND	n___	Tb				
*___	Be	<0.005 ND	n___	Ho		*___	Pt	<0.005 ND	n___	Tl				
*___	Bi	<0.005 ND	*___	In	<0.005 ND	*__	K	<0.005 ND	*___	Th	<0.005 ND			
*___	B	<0.005 ND	n___	Ir		n___	Pr		n___	Tm				
*___	Cd	<0.005 ND	*___	Fe	<0.005 D	n___	Re		*___	Sn	<0.005 ND			
*___	Cs	<0.005 ND	*___	La	<0.005 ND	n___	Rh		*___	Ti	<0.005 ND			
*___	Ca	0.007 D	*___	Pb	<0.005 ND	n___	Rb		n___	W				
n___	Ce		*___	Li	<0.005 ND	n___	Ru		n___	U				
*___	Cr	<0.005 ND	n___	Lu		n___	Sm		s___	V				
*___	Co		*___	Mg	<0.005 ND	n___	Sc		n___	Yb				
*___	Cu	<0.005 ND	*___	Mn	<0.005 ND	*___	Se	<0.005 ND	n___	Y				
n___	Dy		*___	Hg	<0.005 ND	*___	Si	<0.005 ND	*___	Zn	<0.005 ND			
*___	Er	<0.005 ND	*___	Mo	<0.005 ND	*___	Ag	<0.005 ND	n___	Zr				
*___	Eu	<0.005 ND	n___	Nd		*___	Na	<0.005 D						
*___	Gd	<0.005 ND	*___	Ni	<0.005 ND	*___	Sr	<0.005 ND						

* - element checked; i - spectral interference; n – element not checked; D – element detected; ND – element not detected; s - standard element

Storage: Store at Room Temperature (15° to 30°C).

ANAB ACCREDITED ISO 17034 REFERENCE MATERIAL PRODUCER
ISO 17034 Cert No. AR-1936

Produced in accordance with TUV SUD 951215321 registered ISO 9001 Quality Management System

ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY
ISO 17025 Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

Certificate of Analysis

Product Number: ICP-023
Lot Number: 0006684276

Lot Issue Date: 26-May 2022
Expiration Date: 30-Jun 2029

Traceability:
Traceability has been established through an unbroken chain of comparisons, each having stated uncertainties. Comparisons are based on appropriate physical or chemical measurements, including gravimetric or volumetric dilution, where the mass or volume of a solution before and after dilution is measured. The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 819.

Estimation of Uncertainties:
The analyte concentration is reported, with its uncertainty value calculated at the 95% confidence level.

Homogeneity:
This RM was formulated and unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:
This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods and continuing calibration verification.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening and should be processed without delay for the true value to be valid within the stated uncertainties. Do not pipet from the bottle. Do not return any material removed for pipetting to the bottle. Tightly cap the bottle after removing any material and store according to the instructions noted above.

Hazards:
Refer to the Safety Data Sheet for information regarding this RM.

Expiration of Certification:
The certification of this RM is valid, within the measurement uncertainty specified, until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:
The real-time, long term stability of the RM may be monitored over the lifetime of the certification. If substantive changes occur that affect the certification before the expiration of this certificate, Agilent Technologies will notify the purchaser.

Monica Bourgeois
QMS Representative

ISO 17034 Cert No.
AR-1936

Produced in accordance with TUV SUD 951215321
registered ISO 9001 Quality Management System

ISO 17025 Cert No.
AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality