

Certificate of Analysis

Wash-Water Blank

Agilent Part Number: **G1820-60259**

Lot Number: **40-132GST2**

Purity grades:

Matrix:

H₂O: DI Water (CAS No. 7732-18-5) 18Megohm, double deionized (ASTM Type I)

Traceability:

This standard has been produced using ISO 9001 quality procedures.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.01	Cr	<0.01	Ho	<0.01	Nb	<0.01	Ru	<0.01	Th	<0.01
Al	<0.03	Cs	<0.01	In	<0.01	Nd	<0.01	Sb	<0.01	Ti	<0.01
As	<0.01	Cu	<0.01	Ir	<0.01	Ni	<0.01	Sc	<0.01	Tl	<0.01
Au	<0.01	Dy	<0.01	K	<0.1	P	<10	Se	<0.01	Tm	<0.01
B	<0.02	Er	<0.01	La	<0.01	Pb	<0.01	Si	<5	U	<0.01
Ba	<0.01	Eu	<0.01	Li	<0.01	Pd	<0.01	Sm	<0.01	V	<0.01
Be	<0.01	Fe	<0.1	Lu	<0.01	Pr	<0.01	Sn	<0.01	W	<0.01
Bi	<0.01	Ga	<0.01	Mg	<0.01	Pt	<0.01	Sr	<0.01	Y	<0.01
Ca	<0.1	Gd	<0.01	Mn	<0.01	Rb	<0.01	Ta	<0.01	Yb	<0.01
Cd	<0.01	Ge	<0.01	Mo	<0.01	Re	<0.01	Tb	<0.01	Zn	<0.03
Ce	<0.01	Hf	<0.01	Na	<0.03	Rh	<0.01	Te	<0.01	Zr	<0.01
Co	<0.01	Hg	<0.01								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±10% of listed impurities shown above. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: May 31, 2022

Date of expiration: August 31, 2023


 QC Coordinator