



CERTIFICATE OF ANALYSIS

PRODUCT NAME:	AdvanceBio InstantPC G2F / FA2G2 N-Glycan Standard
PRODUCT CODE:	GKPC-305
GLYCAN NAME:	Asialo-, galactosylated biantennary complex N-glycan, core-substituted with fucose
ALTERNATIVE NAMES:	F(6)A2G(4)2, NA2F
LOT NUMBER:	DP22E5039Z
PACK SIZE:	~25 injections (qualitative standard for glycan identification)
PURITY:	≥90% of glycan by UHPLC
FORM:	Dry solid
STORAGE:	Store at –20 °C in the dark before and after reconstitution
EXPIRATION:	August 2026 May be used for 6 months after reconstitution in 100 mM ammonium formate, pH 4.4–5.0, or for 1 month after reconstitution in water.
STRUCTURE ^{1,2,3} :	The glycosylamine form of the glycan is labeled with the fluorescent dye, InstantPC.



Structure Key:

<u>Monosaccharide symbol</u>	<u>Linkage position</u>	<u>Linkage type</u>
 Galactose		— β-linkage
 Mannose		— β-linkage
 Fucose		--- α-linkage
 N-Acetylglucosamine (GlcNAc)		--- α-linkage

Quality Control:




Figure 1 - UHPLC Results: 1 μ L, aqueous of the InstantPC-labeled glycan was injected on a UHPLC System (Agilent 1290 Infinity II) utilizing a 15-minute method under the conditions below:

Time (min)	Flow (mL/min)	% ACN	% Buffer
00.0	1.0	75.0	25.0
12.0	1.0	50.0	50.0
12.1	0.5	40.0	60.0
12.5	0.5	40.0	60.0
12.6	0.5	75.0	25.0
13.0	1.0	75.0	25.0
15.0	1.0	75.0	25.0

Column: Waters ACQUITY UPLC BEH Glycan Column (1.7 μ m, 2.1 x 100 mm)

ACN: Acetonitrile

Buffer: 100 mM ammonium formate, pH 4.4

Flow rate: As stated in table, in ml/min

Temperature: 60 $^{\circ}$ C

Max Pressure: 15,000 psi

Fluorescence Detection: λ_{ex} = 285 nm, λ_{em} = 345 nm

Average Mass⁴: 2049.0

Monoisotopic Mass⁴: 2047.7978

Structural Analysis: The identity of the glycan is confirmed by MALDI-TOF^{5,6} or LC-MS. Agreement was found between the results from mass spectrometry and UHPLC⁷.

Application:

- Qualitative standard for various analytical procedures
- As a migration standard for liquid chromatography

Handling & Reconstitution: The labeled oligo-saccharide is shipped as a dried solid. Use ultra-pure water or an aqueous buffer to dissolve the materials (see Directions for Use for suggested volumes).

Allow the unopened vial to reach ambient temperature and tap on a solid surface to ensure that most of the material is at the bottom of the vial. Gently remove the cap, add the desired volume of ultra-pure water or aqueous buffer, re-cap and mix thoroughly to redissolve all the material.

For maximal recovery, ensure that the cap lining is also rinsed. Centrifuge the reconstituted vial briefly before use.

Make sure that any glassware, plasticware, solvents or reagents used are free of glycosidases and carbohydrate contaminants.

Minimize exposure to elevated temperatures or extremes of pH. Store the reconstituted glycan at -20°C . Allow the vial to equilibrate to ambient temperature before use.

Directions For Use: The amount of InstantPC-labeled glycan standard injected on a UHPLC column is typically $1\ \mu\text{L}$. For our Quality Control testing, one vial was dissolved in $30\ \mu\text{L}$ of water and $1\ \mu\text{L}$ injected on the ACQUITY BEH Glycan column

We suggest reconstituting with $100\ \text{mM}$ ammonium formate, $\text{pH } 4.4\text{--}5.0$ for storage at -20°C for up to 6 months. This buffer is often used as a HILIC mobile phase. Water may also be used for reconstitution, but the recommended storage period is shorter, -20°C for up to 1 month.

For larger injection volumes of InstantPC-labeled glycans ($>1\ \mu\text{L}$), do not use ACN alone to dilute the glycan to match the high organic % at the start of HILIC methods, as this may cause sialylated InstantPC glycans to precipitate. Use 1 part glycan in ammonium formate or water to 3 parts $1:1\ [\text{v/v}]$ ACN:DMF, for a final concentration of 25% aqueous buffer, 37.5% DMF, 37.5% ACN. Dilute only as much as is needed, and freeze the main stock at -20°C . For example, for a $10\ \mu\text{L}$ injection, dilute $5\ \mu\text{L}$ of glycan stock in ammonium formate or water with $15\ \mu\text{L } 1:1\ [\text{v/v}]$ ACN:DMF.

REFERENCES

1. Ceroni A, Maass K, Geyer H, Geyer R, Dell A, Haslam SM. GlycoWorkbench: a tool for the computer-assisted annotation of mass spectra of glycans. *J Proteome Res.* 2008 Apr; 7(4): 1650-9.
2. Harvey DJ, Merry AH, Royle L, Campbell MP, Dwek RA, Rudd PM. Proposal for a standard system for drawing structural diagrams of N- and O-linked carbohydrates and related compounds. *Proteomics* 2009 Aug; 9(15): 3796-801.
3. Harvey DJ, Merry AH, Royle L, Campbell MP, Rudd PM. Symbol nomenclature for representing glycan structures: Extension to cover different carbohydrate types. *Proteomics* 2011 Nov; 11(22): 4291-5.
4. Average mass and monoisotopic mass of the unlabeled glycan (free reducing end form) were calculated using the ExPASy GlycanMass calculator:

<http://web.expasy.org/glycanmass/>

Calculating the Mass of Glycans Labeled with InstantPC.

Mass added to glycan with a free reducing end:
 $\text{Mass}_{\text{Glycan (free reducing end)}} + \text{C}_{14}\text{N}_3\text{O}_2\text{H}_{19} = \text{Mass}_{\text{InstantPC-Labeled Glycan}}$

Mass added by $\text{C}_{14}\text{N}_3\text{O}_2\text{H}_{19}$ (Da):
Monoisotopic: 261.14773
Average: 261.3

Mass added to glycosylamine:
 $\text{Mass}_{\text{Glycan (glycosylamine)}} + \text{C}_{14}\text{N}_2\text{O}_3\text{H}_{18} = \text{Mass}_{\text{InstantPC-Labeled Glycan}}$

Mass added by $\text{C}_{14}\text{N}_2\text{O}_3\text{H}_{18}$ (Da):
Monoisotopic: 262.13174
Average: 262.3
5. James DC, Jenkins N. Analysis of N-glycans by matrix-assisted laser desorption/ionization mass spectrometry; in *A laboratory guide to glycoconjugate analysis*. BioMethods (P. Jackson and J. T. Gallagher, ed) 1997; 9: 91-112.
6. Papac DI, Wong A, Jones AJS. Analysis of acidic oligosaccharides by matrix-assisted laser desorption/ionization time of flight mass spectrometry. *Anal Chem* 1996 Sep 15; 68(18): 3215-3223.
7. Ahn J, Bones J, Yu YQ, Rudd PM, Gilar M. Separation of 2-aminobenzamide labeled glycans using hydrophilic interaction chromatography columns packed with $1.7\ \mu\text{m}$ sorbent. *J Chromatogr B Analyt Technol Biomed Life Sci.* 2010 Feb 1; 878(3-4): 403-8.

Christopher J. Miller Digitally signed by Christopher J. Miller
Date: 2022.08.24 13:04:29 -0700

Authorized Signature