

Reference Material Certificate
Product Information Sheet

Product Name: Organochlorine Pesticide Standard
Product Number: PPM-130-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006703991
Lot Issue Date: 14-Sep-2022
Expiration Date: 30-Apr-2023

Component Name	CERTIFIED VALUES			CAS#	Analyte Lot
	Concentration	Expanded	Uncertainty		
alachlor	100.4	±	0.5 µg/mL	015972-60-8	RM14112
aldrin	100.2	±	0.5 µg/mL	000309-00-2	RM00635
atrazine	100.4	±	0.5 µg/mL	001912-24-9	RM17497
lindane	100.3	±	0.5 µg/mL	000058-89-9	RM02694
dieldrin	100.4	±	0.5 µg/mL	000060-57-1	RM19936
endrin	100.2	±	0.5 µg/mL	000072-20-8	RM14459
heptachlor	100.4	±	0.5 µg/mL	000076-44-8	RM03111
hexachlorobenzene	100.4	±	0.5 µg/mL	000118-74-1	RM03125
hexachlorocyclopentadiene	100.5	±	0.5 µg/mL	000077-47-4	RM12238
methoxychlor	100.2	±	0.5 µg/mL	000072-43-5	RM14186
cis-nonachlor	100.3	±	0.5 µg/mL	005103-73-1	RM03193
trans-nonachlor	100.4	±	0.5 µg/mL	039765-80-5	RM03191
cis-chlordane	100.2	±	0.5 µg/mL	005103-71-9	RM07243
simazine	100.4	±	0.5 µg/mL	000122-34-9	RM00622
trans-chlordane	100.2	±	0.5 µg/mL	005103-74-2	RM02726
heptachlor epoxide - isomer B	100.4	±	0.5 µg/mL	001024-57-3	RM20115

Matrix: methanol (methyl alcohol)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative