



## Results of Analysis

Product:

**OFFGEL pH 3-10 Strip, High Resolution**

Code Number:

**28-9083-93**

Gel Lot No: **10242253**

Customer: Agilent, part no: 5188-6430

Test/Characteristic:	Test method:	Limits:	Results:
<b>1. Function</b>	<b>28-9583-23</b>		
1.1 Relative band position; $R_p$			
- Ferredoxin		0.01 – 0.08	0.04
- $\beta$ -Lactoglobulin, variant A		0.20 – 0.27	0.23
- Carbonic anhydrase		0.53 – 0.60	0.56
- Bovine Hemoglobin, $\beta$ -chain		0.66 – 0.73	0.68
- Bovine Hemoglobin, $\alpha$ -chain		0.82 – 0.89	0.85
1.2 Slope of the gradient Difference between $R_p$ -values for			
- Carbonic anhydrase and Ferredoxin		0.50 – 0.55	0.52
- Bovine Hemoglobin, $\alpha$ -chain and Ferredoxin		0.78 – 0.85	0.82

**Approval date (Year-Month-Day):** 2016-02-29

**Expiry date (Year-Month):**

2017-08

**Manufacturing date(Year-Month):**

2016-02

GE Healthcare Bio-Sciences AB

Björkgatan 30

SE-751 84 Uppsala

Sweden

T + 46 (0)18 612 00 00

F + 46 (0)18 612 12 00

[www.gelifesciences.com/protein-purification](http://www.gelifesciences.com/protein-purification)

Reg.No. SE 55 61 08 1919 01

**Custom Products Lab Media Life Sciences  
Issued (Year-Month-Day) 2016-11-16 by Mikael Söder**

This document has been electronically produced and is valid without a signature.