



Results of Analysis

Product:

OFFGEL pH 4-7 Strip, High Resolution

Code Number:

28-9084-04

Gel Lot No: **10081370**

Customer: **Agilent, part no: 5188-6432**

Test/Characteristic:	Test method:	Limits:	Results:
1. Function	04-0083-62		
1.1 Relative band position; R_p			
- Glycoprotein		0.07 – 0.13	0.11
- β -Lactoglobulin, variant B		0.26 – 0.34	0.31
- Bovin Serum Albumin		0.59 – 0.70	0.65
- Carbonic anhydrase		0.89 – 0.99	0.93
1.2 Slope of the gradient Difference between R_p -values for			
- Bovin Serum Albumin and β -Lactoglobulin, variant B		0.32 – 0.38	0.34
- Carbonic anhydrase and Glycoprotein		0.80 – 0.90	0.82

Approval date (Year-Month-Day): 2012-02-03

Expiry date (Year-Month):

2013-07

Manufacturing date(Year-Month):

2012-01

GE Healthcare Bio-Sciences AB

Björkgatan 30

SE-751 84 Uppsala

Sweden

T + 46 (0)18 612 00 00

F + 46 (0)18 612 12 00

www.gelifesciences.com/protein-purification

Reg.No. SE 55 61 08 1919 01

**Custom Products Lab Media Life Sciences
Issued (Year-Month-Day) 2012-04-16 by Christer Nilsson**

This document has been electronically produced and is valid without a signature.