



Results of Analysis

Product:

OFFGEL pH 4-7 Strip, Low Resolution

Code Number:

28-9084-07

Gel Lot No: 10077656

Customer: Agilent, part no: 5188-6433

Test/Characteristic:	Test method:	Limits:	Results:
1 Function	04-0083-62		
1.1 Relative band position; R_p			
- Glycoprotein		0.06 – 0.14	0.09
- β -Lactoglobulin, variant B		0.25 – 0.35	0.29
- Bovin Serum Albumin		0.58 – 0.71	0.62
- Carbonic anhydrase		0.88 – 1.00	0.91
1.2 Slope of the gradient Difference between R_p -values for			
- Bovin Serum Albumin and β -Lactoglobulin, variant B		0.31 – 0.39	0.33
- Carbonic anhydrase and Glycoprotein		0.79 – 0.91	0.82

Approval date (Year-Month-Day): 2012-01-19

Expiry date (Year-Month):

2013-07

Manufacturing date(Year-Month):

2012-01

GE Healthcare Bio-Sciences AB

Björkgatan 30

SE-751 84 Uppsala

Sweden

T + 46 (0)18 612 00 00

F + 46 (0)18 612 12 00

www.gelifesciences.com/protein-purification

Reg.No. SE 55 61 08 1919 01

**Custom Products Lab Media Life Sciences
Issued (Year-Month-Day) 2012-06-19 by Christer Nilsson**

This document has been electronically produced and is valid without a signature.