



Results of Analysis

Product:**OFFGEL Buffer pH 3-10, 1 ml****Code Number:****28-9090-32****Gel Lot No: 10044450**

Test/Characteristic:	Test method:	Limits:	Results:
1. pH	45-0312-71	6.70 – 7.20	7.10
2.. Buffering capacity; mmol·pH ⁻¹ ·ml ⁻¹	45-0310-37	0.3 – 0.6	0.4

Date of analysis: 2010-08-23**Expiry:**

2013-08

Manufacturing date:2010-08

GE Healthcare Bio-Sciences AB

Björkgatan 30

SE-751 84 Uppsala

Sweden

T + 46 (0)18 612 00 00

F + 46 (0)18 612 12 00

www.gelifesciences.com/protein-purification

Reg.No. SE 55 61 08 1919 01

**Custom Products Life Sciences
Issued 2010-08-31 by Christer Nilsson**

This document has been electronically produced and is valid without a signature.