



Results of Analysis

Product: OFFGEL Buffer pH 3-10, 1 ml
Code Number: 28-9090-32

Gel Lot No: 10185576

Test/Characteristic:	Test method:	Limits:	Results:
1. pH	45-0312-71	6.70 – 7.20	7.12
2.. Buffering capacity; mmol·pH ⁻¹ ·ml ⁻¹	45-0310-37	0.3 – 0.6	0.4

Date of analysis: 2012-06-05

Expiry: 2016-06
Manufacturing date: 2013-06

GE Healthcare Bio-Sciences AB
Björkgatan 30
SE-751 84 Uppsala
Sweden
T + 46 (0)18 612 00 00
F + 46 (0)18 612 12 00
www.gelifesciences.com/protein-purification
Reg.No. SE 55 61 08 1919 01

Custom Products Life Sciences
Issued 2013-06-13 by Christer Nilsson

This document has been electronically produced and is valid without a signature.