



Results of Analysis

Product:	Code Number:
OFFGEL Buffer pH 3-10, 1 ml	28-9090-32
Gel Lot No: 10241212	

Test/Characteristic:	Test method:	Limits:	Results:
1. pH	45-0312-71	6.70 – 7.20	7.0
2.. Buffering capacity; mmol·pH ⁻¹ ·ml ⁻¹	45-0310-37	0.3 – 0.6	0.4

Date of analysis: 2015-08-27

Expiry:

2018-08

Manufacturing date:

2015-08

GE Healthcare Bio-Sciences AB

Björkgatan 30

SE-751 84 Uppsala

Sweden

T + 46 (0)18 612 00 00

F + 46 (0)18 612 12 00

www.gelifesciences.com/protein-purification

Reg.No. SE 55 61 08 1919 01

Custom Products Life Sciences
Issued 2015-12-02 by Mikael Söder

This document has been electronically produced and is valid without a signature.