



Results of Analysis

Product:				Code Number:
OFFGEL Buffer pH 3-10, 1 ml				28-9090-32
Gel Lot No: 10251401				

Test/Characteristic:	Test method:	Limits:	Results:
1. pH	45-0312-71	6.7 – 7.2	6.7
2.. Buffering capacity; mmol·pH ⁻¹ ·ml ⁻¹	45-0310-37	0.3 – 0.6	0.4

Date of analysis: 2016-06-29

Expiry:

2019-06

Manufacturing date:

2016-06

GE Healthcare Bio-Sciences AB
Björkgatan 30
SE-751 84 Uppsala
Sweden
T + 46 (0)18 612 00 00
F + 46 (0)18 612 12 00
www.gelifesciences.com/protein-purification
Reg.No. SE 55 61 08 1919 01

Custom Products Life Sciences
Issued 2016-11-16 by Mikael Söder

This document has been electronically produced and is valid without a signature.