



Results of Analysis

Product:	Code Number:
OFFGEL Buffer pH 4-7, 1 ml	28-9090-35
Lot number: 10029342	
Customer: Agilent, part no: 5188-6438	

Test/Characteristic:	Test method:	Limits:	Results:
1. pH	45-0312-71	5.25 – 5.85	5.61
2. Buffering Capacity; mmol·pH ⁻¹ ·ml ⁻¹	45-0310-37	0.3 – 0.6	0.4

Date of analysis: 2009-01-08	Expiry date: 2012-01
	Manufacturing date: 2009-01

GE Healthcare Bio-Sciences AB
Björkgatan 30
SE-751 84 Uppsala
Sweden
T + 46 (0)18 612 00 00
F + 46 (0)18 612 12 00
www.gehealthcare.com
Reg.No. SE 55 61 08 1919 01

Issuing department: Custom Products Life Sciences
Issuing date: 2009-01-23
Issued by: Christer Nilsson

This document has been electronically produced and is valid without a signature.