

Certificate of Analysis

Caffeine Standards Kit for Capillary LC

Part No 5065-4420

Lot No HC015698

Concentrations:

Caffeine in water:

Nominal value	Effective value determined by UV-spectroscopy using the absorption maximum at 273 nm
200 µg/ml	200.6 µg/ml + 2 %
100 µg/ml	99.6 µg/ml + 2 %
20 µg/ml	19.7 µg/ml + 2 %
4 µg/ml	4.0 µg/ml + 3 %
2 µg/ml	2.0 µg/ml + 5 %

Purity grades:

Caffeine: extra pure, acc. to Ph. Eur.
Concentration (acid. titration): 98.5 – 101.5 %
Heavy metals (as Pb): < 0.001%
Loss on drying (105°C): < 0.5%


Water: 18.2 MΩ x cm

The standards in this kit have been produced gravimetrically under controlled environmental conditions. Balances used are calibrated regularly against PTB (Physikalisch Technische Bundesanstalt [Federal Physical-Technical Institute] – Braunschweig – Germany) traceable weight sets. The manufacturing, testing and distribution of these standards meet the requirements of the quality management norm DIN EN ISO 9001:2008.

The standards have been analysed on a high-performance UV/VIS/NIR spectrophotometer. The spectrophotometer is regularly validated for accuracy and reproducibility of absorbance and wavelength as well as for linearity, baseline drift, stray light and spectral resolution power using the following testing materials:

Absorbance: NIST SRM 1930 and double aperture method
Wavelength: Intrinsic values of a holmium oxide solution Merck reference material Cat. No. 108166 [1], emission lines of D₂ and Hg lamps
Stray light: Merck Cat. No. 108164 (KCl standard, 200 nm), Merck Cat. No. 108163 (NaI standard, 220 nm), Merck Cat. No. 108161 (NaNO₂ standard, 370 nm)
Spectral resolution power: Half width value of D₂ emission lines for checking the effective optical bandwidth

UV/VIS Spectrum Caffeine in water [concentration: 100 µg/ml]



[1] J. C. Travis et al., J. Phys. Chem. Ref. Data (2005), 34(1), 41-56.

Date of release: 27.05.2010

Expiration date: 31.05.2013

Dr. Stefan Frey
(responsible laboratory manager quality control)