



Certificate of Analysis

ICP-MS Stock Tuning Solution (100 mL)

Agilent Part Number: 5188-6564

Lot Number: 8-224MFY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.0 µg/mL	3110*	Ce ₂ (CO ₃) ₃ ·x(H ₂ O)	99.99+	Tl	7440-28-0	10.0 µg/mL	9.96 µg/mL	3158*	TlNO ₃	99.99+
Co	7440-48-4	10.0 µg/mL	9.98 µg/mL	3113*	Co	99.99+	Y	7440-65-5	10.0 µg/mL	9.99 µg/mL	3167a*	Y ₂ O ₃	99.99+
Li	7439-93-2	10.0 µg/mL	9.98 µg/mL	3129a*	Li ₂ CO ₃	99.99+							

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 8-140MF, 7-246MF, 3-210SG.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.3	Cs	<0.04	Ho	<0.01	Nd	<0.01	Ru	0.1	Te	<0.2
Al	1	Cu	<0.1	In	<0.04	Ni	<0.01	Sb	0.08	Th	<0.01
As	<1	Dy	<0.01	Ir	<0.02	P	<100	Sc	<0.1	Ti	<0.5
Au	<0.06	Er	<0.01	K	<80	Pb	0.2	Se	<0.5	Tm	<0.01
B	<0.4	Eu	<0.01	La	<0.01	Pd	<0.5	Si	<100	U	<0.01
Ba	0.2	Fe	<1	Lu	<0.01	Pr	0.2	Sm	<0.01	V	<0.1
Be	<0.1	Ga	0.5	Mg	0.7	Pt	<0.02	Sn	<0.1	W	<0.01
Bi	<0.01	Gd	0.2	Mn	<0.2	Rb	<0.3	Sr	<0.2	Yb	<0.01
Ca	5	Ge	<0.04	Mo	0.4	Re	<0.02	Ta	<0.01	Zn	0.4
Cd	0.08	Hf	<0.01	Na	2	Rh	<0.01	Tb	1	Zr	<0.1
Cr	<0.3	Hg	<0.2	Nb	<0.02						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: March 15, 2023

Date of expiration: September 30, 2024

Brijender S. Tonk
 QC Coordinator