Certificate of Analysis

Electron Capture Detector Sample

Agilent Part Number: 5183-0379 **Sample Lot Number:** S0227

This ULTRAstandard (TM) solution was gravimetrically prepared, and concentration was confirmed analytically. All operations were carried out under ULTRA Scientific's ISO 9001 registered quality system. The true value for each analyte, determined gravimetrically, is listed below with an uncertainty level of +/- 0.5%. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Concentrations:

gamma-BHC	0.033	mg/L


Solvent: Isooctane

Purity grades:

gamma-BHC	99%
Isooctane	99%

Analytical Spectrum or Chromatogram

LC Chromatogram – gamma-BHC in isooctane



Date of release: 13 February 2001
Date of expiration: 31 March 2003

Michael Consiglio
Quality Control Manager

Agilent Technologies
Innovating the HP Way

Manufactured for Agilent by ULTRA Scientific, Inc., USA