

# Certificate of Analysis

## ICP-MS Stock Tuning Solution (100 mL)

**Agilent Part Number: 5188-6564**
**Lot Number: 23-29VYY2**

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	9.92 µg/mL	3110*	Ce <sub>2</sub> (CO <sub>3</sub> ) <sub>3</sub> ·x(H <sub>2</sub> O)	99.99+	Tl	7440-28-0	10.0 µg/mL	9.92 µg/mL	3158*	TlNO <sub>3</sub>	99.99+
Co	7440-48-4	10.0 µg/mL	9.93 µg/mL	3113*	Co	99.99+	Y	7440-65-5	10.0 µg/mL	10.1 µg/mL	3167a*	Y <sub>2</sub> O <sub>3</sub>	99.99+
Li	7439-93-2	10.0 µg/mL	9.90 µg/mL	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+							

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

**Purity grades:**

Starting Materials: Shown above

Matrix:

 2% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 22-85VY, 22-150VY.

**Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:**


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.1	Cs	<0.01	Ho	<0.01	Nd	<0.03	Ru	<0.3	Te	<0.6
Al	0.6	Cu	<0.1	In	<0.03	Ni	0.1	Sb	<0.07	Th	<0.01
As	<0.3	Dy	<0.01	Ir	<0.03	P	<100	Sc	<0.2	Ti	<0.3
Au	<0.2	Er	<0.01	K	<2	Pb	<0.04	Se	<0.6	Tm	<0.01
B	<0.9	Eu	0.06	La	<0.1	Pd	<0.5	Si	<100	U	<0.01
Ba	<0.3	Fe	<2	Lu	<0.01	Pr	0.1	Sm	<0.05	V	<0.04
Be	<0.3	Ga	0.4	Mg	<0.1	Pt	<0.03	Sn	<0.09	W	<0.4
Bi	<0.03	Gd	<0.1	Mn	<0.4	Rb	<0.05	Sr	<0.09	Yb	<0.01
Ca	<8	Ge	<0.6	Mo	<0.2	Re	<0.02	Ta	<0.2	Zn	<0.4
Cd	<0.08	Hf	<0.01	Na	0.5	Rh	<0.04	Tb	0.7	Zr	<0.1
Cr	0.3	Hg	<0.2	Nb	<0.2						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

**Date of release:** April 30, 2016

**Date of expiration:** October 31, 2017


  
 QC Coordinator