



Certificate of Analysis

ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465

Lot Number: 2-57KCY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.1 µg/mL	3110*	Ce ₂ (CO ₃) ₃ ·x(H ₂ O)	99.99+	Mg	7439-95-4	10.0 µg/mL	10.1 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	10.1 µg/mL	3113*	Co	99.99+	Tl	7440-28-0	10.0 µg/mL	9.98 µg/mL	3158*	TlNO ₃	99.99+
Li	7439-93-2	10.0 µg/mL	10.1 µg/mL	3129a*	Li ₂ CO ₃	99.99+	Y	7440-65-5	10.0 µg/mL	10.0 µg/mL	3167a*	Y ₂ O ₃	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 4-127MKB, 1-235YJ, 3-150MKB.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.2	Cs	<0.02	Ho	<0.1	Nd	<0.01	Ru	<0.06	Te	<0.4
Al	3	Cu	<0.3	In	<0.01	Ni	0.3	Sb	<0.5	Th	<0.04
As	0.2	Dy	<0.01	Ir	<0.1	P	<100	Sc	<0.1	Ti	<0.3
Au	<0.04	Er	<0.01	K	0.5	Pb	<0.07	Se	<2	Tm	<0.01
B	<2	Eu	<0.01	La	0.06	Pd	<0.1	Si	<100	U	<0.01
Ba	<0.2	Fe	2	Lu	<0.01	Pr	0.1	Sm	<0.01	V	<0.5
Be	<0.1	Ga	<0.5	Mn	0.2	Pt	<0.01	Sn	<0.07	W	<0.6
Bi	<0.01	Gd	<0.2	Mo	<0.5	Rb	<0.06	Sr	<0.05	Yb	<0.01
Ca	3	Ge	<0.2	Na	0.6	Re	<0.01	Ta	<0.1	Zn	1
Cd	<0.03	Hf	<0.01	Nb	<0.2	Rh	<0.04	Tb	<0.4	Zr	<0.1
Cr	0.3	Hg	<0.1								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: June 15, 2019

Date of expiration: December 31, 2020

Brijender S Tonk
 QC Coordinator