



# Certificate of Analysis

## ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465

Lot Number: 2-60YJY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.0 µg/mL	3110*	Ce <sub>2</sub> (CO <sub>3</sub> ) <sub>3</sub> ·x(H <sub>2</sub> O)	99.99+	Mg	7439-95-4	10.0 µg/mL	9.99 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	9.93 µg/mL	3113*	Co	99.99+	Tl	7440-28-0	10.0 µg/mL	10.0 µg/mL	3158*	TlNO <sub>3</sub>	99.99+
Li	7439-93-2	10.0 µg/mL	9.94 µg/mL	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+	Y	7440-65-5	10.0 µg/mL	9.96 µg/mL	3167a*	Y(NO <sub>3</sub> ) <sub>3</sub> ·6H <sub>2</sub> O	99.99+

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

### Purity grades:

Starting Materials: Shown above

Matrix:

2% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

### Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 1-201YJ, 1-135YJ.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.2	Cs	<0.04	Ho	0.04	Nd	<0.02	Ru	<0.06	Te	<0.6
Al	1	Cu	<0.3	In	<0.02	Ni	<0.2	Sb	<0.07	Th	<0.02
As	<0.3	Dy	0.05	Ir	<0.09	P	<100	Sc	<0.2	Ti	<0.2
Au	<0.2	Er	<0.01	K	0.4	Pb	<0.03	Se	<10	Tm	<0.01
B	<3	Eu	<0.01	La	0.07	Pd	<0.4	Si	<100	U	<0.01
Ba	<0.08	Fe	2	Lu	<0.02	Pr	0.2	Sm	<0.02	V	<0.2
Be	<0.3	Ga	0.4	Mn	0.4	Pt	<0.2	Sn	<0.04	W	<0.4
Bi	<0.05	Gd	0.2	Mo	<0.2	Rb	<0.09	Sr	<0.04	Yb	<0.01
Ca	4	Ge	<0.09	Na	<2	Re	<0.01	Ta	<0.2	Zn	0.6
Cd	<0.05	Hf	<0.01	Nb	<0.07	Rh	<0.2	Tb	0.5	Zr	<0.4
Cr	0.6	Hg	<0.2								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: July 31, 2019

Date of expiration: January 31, 2021

*Brijender S. Tank*  
QC Coordinator