



CERTIFICATE OF ANALYSIS

Agilent Product Name: A-SOLV ICP Solvent

Agilent Part No: 5190-8717

Lot No: 7110656269

Product Specifications

Matrix: Pure, Kerosene type distillate

Intended Use: A-SOLV ICP Solvent is a proprietary solvent that is intended for use as a calibration blank solution or diluent for the analysis of petroleum products or other organic matrices using inductively coupled plasma optical emission spectroscopy (ICP-OES), and inductively coupled plasma mass spectrometry (ICP-MS).

Certification & Traceability: This solution was prepared under a quality management system that is accredited to **ISO/IEC 17025**, and registered to **ISO 9001**. It was analyzed for trace metals by inductively coupled plasma optical emission spectroscopy (ICP-OES), with traceability to NIST SRM 1085b.

Trace Concentrations (µg/g)

Ag	<0.25	Cr	<0.25	Mo	<0.25	Si	<0.25
Al	<0.25	Cu	<0.25	Na	<0.25	Sn	<0.25
As	<0.25	Fe	<0.25	Ni	<0.25	Sr	<0.25
B	<0.25	Hg	<0.25	P	<0.25	Ti	<0.25
Ba	<0.25	K	<0.25	Pb	<0.25	Tl	<0.25
Bi	<0.25	La	<0.25	S	<1	V	<0.25
Ca	<0.25	Li	<0.25	Sb	<0.25	Y	<0.25
Cd	<0.25	Mg	<0.25	Sc	<0.25	Zn	<0.25
Co	<0.25	Mn	<0.25	Se	<0.25		

Instructions for Use: Agilent Technologies recommends that the solution be thoroughly mixed by repeated shaking or swirling of the bottle immediately prior to use. The solution should be kept tightly capped. Store at controlled room temperature per USP 35 (10.30.60). Do not freeze, heat, or expose to direct sunlight. Minimize exposure to moisture or high humidity.

Period of Validity: Agilent Technologies ensures the accuracy of this solution until the expiration date shown below, provided the instructions for use are followed. During the period of validity, the purchaser will be notified if this product is recalled due to any significant changes in the stability of the solution.

Sample lot approver:

Chuck Goudreau, Certifying Officer

Date of release: 28 January 2019

Date of expiration: 18 July 2020

Hazard Information: Refer to the Safety Data Sheet (SDS), which can be obtained at www.agilent.com/chem/sds.

Homogeneity: This solution was determined to be homogeneous by procedures consistent with the requirements of ISO Guide 34 and ISO Guide 35. Replicate samples of the finished solution were analyzed to confirm its homogeneity, in accordance with QSP 6-13 Assessment of Homogeneity and Stability. To ensure homogeneity, users should not take a smaller sub-sample than specified in the Instructions for Use, as doing so will invalidate the certified values and uncertainties.

Further Information: Please contact Agilent Technologies for further information about this CRM.

Quality Certifications: This CRM was prepared under a quality management system that is:

- Registered to ISO 9001 – Quality Management Systems – Requirements (TUV NORD Cert. No. 44 100 16560231)
- Accredited to ISO Guide 34 – General Requirements for the Competence of Reference Material Producers (A2LA Cert. No. 2848.02)
 - ISO Guide 34 references additional requirements specified in ISO Guide 31 and ISO Guide 35.
- Accredited to ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories (A2LA Cert. No. 2848.01)