Agilent Technologies

SUPELCO

Certificate of Analysis

PFTBA 100g Bulk


Agilent Part Number: 8500-4394 Sample Lot Number: LB05799

Concentration:
GCMS Neat Material: Perfluorotributylamine CAS # 311-89-7 $(CF_3(CF_2)_3)_3N$

Identification:
Capillary GC/MS was used to determine the purity grade of perfluorotributylamine, 99.9%

Traceability:
This standard has been produced gravimetrically using ISO9001 quality procedures. Balances used are calibrated regularly and meet the requirements of ISO. High-resolution gas chromatography mass spectroscopy was used to verify the standard identity.

Analytical Spectrum or Chromatogram: Fragmentation Pattern



Date of Release: 19 July 2002
Date of Expiration: 19 July 2004

Elwood Doughty
QA Manager
Supelco, Inc.