

Certificate of Analysis

Product Name: Pesticide Standard 4

Product Number: AGI-P03-4

Lot Issue Date: 11-May-2023

Lot Number: 0006742773

Expiration Date: 31-Dec-2024

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
bromacil	000314-40-9	RM19730	100.5 ± 0.5 µg/mL
cycloate	001134-23-2	RM16258	100.2 ± 0.5 µg/mL
endrin	000072-20-8	RM14459	99.9 ± 0.5 µg/mL
epoxiconazole	133855-98-8	RM13872	100.0 ± 0.5 µg/mL
fenoithiocarb	062850-32-2	RM17099	100.4 ± 0.5 µg/mL
hexachlorobenzene	000118-74-1	RM03125	100.1 ± 0.5 µg/mL
iprobefos	026087-47-8	RM17505	100.5 ± 0.5 µg/mL
dicofol	000115-32-2	RM14743	100.5 ± 0.5 µg/mL
quinoxiphen	124495-18-7	RM14287	100.4 ± 0.5 µg/mL
propazine	000139-40-2	RM14562	100.2 ± 0.5 µg/mL
tebupirimphos	096182-53-5	RM12654-01	99.6 ± 0.5 µg/mL
tetradifon	000116-29-0	RM16748	100.3 ± 0.5 µg/mL
thiobencarb	028249-77-6	RM16858	100.5 ± 0.5 µg/mL
triadimenol	055219-65-3	RM16846	100.2 ± 0.5 µg/mL

Matrix: acetonitrile

Storage Conditions: Store Refrigerated (2° to 8°C)



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 951215321

Page: 1 of 2

CSD-QA-015.2



ISO 17025 Cert
No. AT-1937



Certificate of Analysis

Product Number: AGI-P03-4

Lot Number: 0006742773

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 951215321

Page: 2 of 2

CSD-QA-015.2



ISO 17025 Cert
No. AT-1937