



Certificate of Analysis

Acrolein-Acrylonitrile Mixture

Product Number: AMN-623

Page: 1 of 1

Lot Number: CM-4013

Lot Issue Date: 29-Jul-2015

Expiration Date: 30-Nov-2015

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
acrolein	000107-02-8	RM08225	2010 ± 10 µg/mL
acrylonitrile	000107-13-1	RM08217	2010 ± 10 µg/mL


Matrix: methanol (methyl alcohol)

Storage: Store at < 4°C.

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


William J. Leary
Quality Assurance Manager