Agilent Technologies

Fluka Analytical

Certificate of Analysis

Description : Amino Acid Standard (0.1 nmol/µl) 10/pk

Part.-No. : 5061-3332 Production date : 04 Dec 2013

Lot-No. : BCBM3214V (99414) Expiration date : 04 Dec 2015

No.	Component	Molecular Weight	Final Concentration (mg/ml)
01	**L-Alanine**	89.10	0.00891
02	**L-Arginine**	174.20	0.01742
03	**L-Aspartic Acid**	133.11	0.01331
04	**L-Cystine**	240.30	0.02403
05	**L-Glutamic Acid**	147.13	0.01471
06	**Glycine**	75.07	0.00751
07	**L-Histidine Hydrochloride Monohydrate**	209.63	0.02096
08	**L-Isoleucine**	131.18	0.01312
09	**L-Leucine**	131.18	0.01312
10	**L-Lysine Hydrochloride**	182.65	0.01827
11	**L-Methionine**	149.21	0.01492
12	**L-Phenylalanine**	165.19	0.01652
13	**L-Proline**	115.13	0.01151
14	**L-Serine**	105.09	0.01051
15	**L-Threonine**	119.12	0.01191
16	**L-Tyrosine**	181.19	0.01812
17	**L-Valine**	117.15	0.01172

Purity Determination:
Amino acids by Titration, Loss on drying, Residue on ignition, IR-Spectroscopy, MicroSelect-Test, Trace Metal and Trace Anion Determination
Hydrochloric acid by Titration, MicroSelect-Test, Trace Metal and Trace Anion Determination

Raw materials
All raw materials used to prepare this amino acid standard are of the highest available purity (> 99%) and are routinely analyzed according to the above-mentioned purity-determinations.

Manufacturing
We employ precise measuring techniques in manufacturing this amino acid standard. Mass is determined with electronic balances capable of weighing to 0.0001 g and calibrated by the Swiss Office of Weights and Measures. Volume ist determined in dedicated high-purity borosilicate volumetric flasks capable of measuring a 2000 ml-volume with an accuracy of 0.3%.

Packaging and Storage
The final solution is handled under argon, filled into 1 ml amber ampoules under inert gas (argon) and sealed. 10 finished ampoules are packaged into a set and stored at 4 °C.

Stability
Every individual lot of the product is subjected to reanalysis and the experience allows to set the shelf life to two years, if the product is stored as received at 4 °C. The guaranteed stability is not applicable to ampoules stored after opening, even if resealed.

Analytical Quality Control
The scope of the analytical testing procedures covers identity, purity, homogeneity, accuracy, function test in amino acid analysis and stability of the finished product

Description	Lot Analysis	Specifications
Aspect	clear, colorless liquid	clear, colorless liquid
Density (20/4)	1.000	1.00 ± 0.01 g/ml
Indes of Refraction (20/D)	1.334	1.334 ± 0.005
Amino Acid Analysis :	corresponds	corresponds
- Identity	corresponds	corresponds
- Purity	corresponds	corresponds
- Concentration Accuracy	corresponds	± 2.5% relative to primary standard

Buchs, 04.12.2013

Sigma-Aldrich Production GmbH
Industriestrasse 25
CH-9471 Buchs / Switzerland

A. Schneider

Dr. A. Schneider, Analytical Product Manager
Quality Control, Buchs / Switzerland

Sigma-Aldrich warrants that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product for its particular use. See invoice for additional terms and conditions of sale.