



Certificate of Analysis

Pesticides Matrix Spiking Solution

Product Number: CLP-202-1

Page: 1 of 1

Lot Number: CR-2288

Lot Issue Date: 23-May-2017

Expiration Date: 30-Jun-2018

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
lindane	000058-89-9	RM02694	99.9 ± 0.5 µg/mL
heptachlor	000076-44-8	RM03111	99.9 ± 0.5 µg/mL
aldrin	000309-00-2	RM00635	99.9 ± 0.5 µg/mL
dieldrin	000060-57-1	RM03957	249.9 ± 1.2 µg/mL
endrin	000072-20-8	RM00557	250.2 ± 1.3 µg/mL
4,4'-DDT	000050-29-3	RM06888	249.5 ± 1.2 µg/mL

Matrix: acetone


Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Monica Bourgeois
Director of QA/RA