



Certificate of Analysis

Pesticides Matrix Spiking Solution

Product Number: CLP-209X

Page: 1 of 1

Lot Number: CS-0881

Lot Issue Date: 23-Feb-2018

Expiration Date: 31-Mar-2020

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	0.501 ± 0.003 µg/mL
gamma-BHC	000058-89-9	RM02694	0.500 ± 0.003 µg/mL
dieldrin	000060-57-1	RM03957	1.00 ± 0.01 µg/mL
4,4'-DDT	000050-29-3	RM06888	1.00 ± 0.01 µg/mL
endrin	000072-20-8	RM00557	1.00 ± 0.01 µg/mL
heptachlor	000076-44-8	RM03111	0.500 ± 0.003 µg/mL

Matrix: methanol (methyl alcohol)


Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Monica Bourgeois
Director of QA/RA