



Certificate of Analysis

Pesticides Individual Standard Mixture A

Product Number: CLP-216D

Page: 1 of 1

Lot Number: CH-2653

Lot Issue Date: 15-Aug-2011

Expiration Date: 30-Sep-2014

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
endosulfan I	000959-98-8	ER012105-02	20.0 ± 0.1 ng/mL
endrin	000072-20-8	256-27A	40.0 ± 0.2 ng/mL
4,4'-DDD	000072-54-8	W800	40.0 ± 0.2 ng/mL
4,4'-DDT	000050-29-3	02230KY	40.0 ± 0.2 ng/mL
heptachlor	000076-44-8	33379-48	20.0 ± 0.1 ng/mL
methoxychlor	000072-43-5	380-133B	200.0 ± 1.0 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	360-147A	20.1 ± 0.1 ng/mL
alpha-BHC	000319-84-6	27275-15	20.0 ± 0.1 ng/mL
gamma-BHC	000058-89-9	NT01899	20.0 ± 0.1 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM00410	40.0 ± 0.2 ng/mL
dieldrin	000060-57-1	ER030105-03	40.0 ± 0.2 ng/mL

Matrix: hexane

Storage: Store at Room Temperature (18-25° C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager