



Certificate of Analysis

Pesticides Individual Standard Mixture A

Product Number: CLP-216E

Page: 1 of 1

Lot Number: CL-1651Z

Lot Issue Date: 17-Jul-2017

Expiration Date: 31-Aug-2020

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alpha-BHC	000319-84-6	RM042725-02	80.0 ± 0.4 ng/mL
gamma-BHC	000058-89-9	RM02694-04	80.0 ± 0.4 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM00348-02	160.0 ± 0.8 ng/mL
dieldrin	000060-57-1	RM03957-01	160.0 ± 0.8 ng/mL
endosulfan I	000959-98-8	RM00834-16	80.0 ± 0.4 ng/mL
endrin	000072-20-8	RM00557-02	160.0 ± 0.8 ng/mL
4,4'-DDD	000072-54-8	RM00615-01	160.0 ± 0.8 ng/mL
4,4'-DDT	000050-29-3	RM02881-01	160.0 ± 0.8 ng/mL
heptachlor	000076-44-8	RM03111-02	80.0 ± 0.4 ng/mL
methoxychlor	000072-43-5	RM04555-08	803.2 ± 4.0 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM03890-02	80.0 ± 0.4 ng/mL

Matrix: hexane


Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Monica Bourgeois
Director of QA/RA