

Pesticides Individual Standard Mixture B

Product Number: CLP-226C**Page:** 1 of 1**Lot Number:** CH-0079**Lot Issue Date:** 01-Feb-2011**Expiration Date:** 28-Feb-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	ER012606-03	5.00 ± 0.03 ng/mL
beta-BHC	000319-85-7	ER052506-01	5.00 ± 0.03 ng/mL
delta-BHC	000319-86-8	ER051805-01	5.00 ± 0.03 ng/mL
alpha-chlordane	005103-71-9	27455-76	5.00 ± 0.03 ng/mL
gamma-chlordane	005103-74-2	ER061906-04	5.00 ± 0.03 ng/mL
4,4'-DDE	000072-55-9	X102	10.0 ± 0.1 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM00410	10.0 ± 0.1 ng/mL
endosulfan sulfate	001031-07-8	32455-46	10.0 ± 0.1 ng/mL
endrin aldehyde	007421-93-4	34701-21	10.0 ± 0.1 ng/mL
endrin ketone	053494-70-5	32455-08	10.0 ± 0.1 ng/mL
endosulfan II	033213-65-9	ER031704-03	10.0 ± 0.1 ng/mL
heptachlor epoxide - isomer B	001024-57-3	ER090105-05	5.00 ± 0.03 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	360-147A	5.00 ± 0.03 ng/mL

Matrix: hexane**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager