



Certificate of Analysis

Pesticides Evaluation Standards Mixture

Product Number: CLP-231

Page: 1 of 1

Lot Number: CM-4429

Lot Issue Date: 20-Aug-2015

Expiration Date: 30-Sep-2017

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	0.999 ± 0.005 µg/mL
endrin	000072-20-8	RM00557	1.99 ± 0.01 µg/mL
4,4'-DDT	000050-29-3	RM06888	2.00 ± 0.01 µg/mL
dibutyl chlorendate	001770-80-5	NT01609	1.00 ± 0.01 µg/mL


Matrix: isooctane (2,2,4-trimethylpentane)

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality