

Pesticides Performance Evaluation Mixture

Product Number CLP-250

Page: 1 of 1

Lot Number: CM-2452

Lot Issue Date: 14-May-2015

Expiration Date: 30-Jun-2018

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alpha-BHC	000319-84-6	RM10237	10.0 ± 0.1 µg/mL
beta-BHC	000319-85-7	RM09584	10.0 ± 0.1 µg/mL
gamma-BHC	000058-89-9	RM02694	10.0 ± 0.1 µg/mL
endrin	000072-20-8	RM00557	50.1 ± 0.3 µg/mL
4,4'-DDT	000050-29-3	RM06888	100.4 ± 0.5 µg/mL
methoxychlor	000072-43-5	RM09243	250.9 ± 1.3 µg/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM04082	20.1 ± 0.1 µg/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM01256	20.0 ± 0.1 µg/mL

Matrix: hexane**Storage:** Store at Room Temperature (15-30°C). Do Not Refrigerate.

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.