

## Custom Standard

Product Number: CUS-10970

Page: 1 of 1

Lot Number: CG-3922

Lot Issue Date: Nov-2010

Expiration Date: Dec-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
phenol	000108-95-2	1373574	1002 ± 5 ng/μL
2-methylphenol	000095-48-7	09416AA	1002 ± 5 ng/μL
3-methylphenol	000108-39-4	03517AG	1002 ± 5 ng/μL
4-methylphenol	000106-44-5	09410PI	1004 ± 5 ng/μL
2,3-dimethylphenol	000526-75-0	09228MD	1000 ± 5 ng/μL
2,5-dimethylphenol	000095-87-4	18819BQ	1002 ± 5 ng/μL
2,6-dimethylphenol	000576-26-1	00718TX	1002 ± 5 ng/μL

**Matrix:** methylene chloride (dichloromethane)**Storage:** Store at Room Temperature (18-25° C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851-01

ISO 9001:2008  
Registered  
TUV USA, Inc.  
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA  
401-294-9400 Fax: 295-2330  
[www.ultrasci.com](http://www.ultrasci.com)

  
William J. Leary  
Quality Assurance Manager