

## Custom Standard

Product Number: CUS-11098

Page: 1 of 1

Lot Number: CG-3702

Lot Issue Date: Oct-2010

Expiration Date: Nov-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
benzene	000071-43-2	34097	1003 ± 5 µg/mL
toluene	000108-88-3	47144	1003 ± 5 µg/mL
styrene	000100-42-5	LR-11228MQ	1003 ± 5 µg/mL
o-xylene	000095-47-6	DO-06834CO	1003 ± 5 µg/mL
ethylbenzene	000100-41-4	01558CC	1003 ± 5 µg/mL
m-xylene	000108-38-3	DI-00459CI	1003 ± 5 µg/mL
p-xylene	000106-42-3	03747LN	1003 ± 5 µg/mL
isopropylbenzene	000098-82-8	02827BH	1003 ± 5 µg/mL
n-propylbenzene	000103-65-1	LO-14503MR	1003 ± 5 µg/mL
3-ethyltoluene	000620-14-4	10408EA	1003 ± 5 µg/mL
4-ethyltoluene	000622-96-8	1180977	1003 ± 5 µg/mL
1,3,5-trimethylbenzene	000108-67-8	KM-02011HM	1003 ± 5 µg/mL
2-ethyltoluene	000611-14-3	04524BH	1003 ± 5 µg/mL
1,2,4-trimethylbenzene	000095-63-6	BO-13528BI	1003 ± 5 µg/mL
1,2,3-trimethylbenzene	000526-73-8	395-89B	1003 ± 5 µg/mL
4-isopropyltoluene	000099-87-6	PP-05104CP	1002 ± 5 µg/mL
n-butylbenzene	000104-51-8	AA-28519CO	1002 ± 5 µg/mL

**Matrix:** methanol (methyl alcohol)**Storage:** Store at < 4° C


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851-01

ISO 9001:2008  
Registered  
TUV USA, Inc.  
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA  
401-294-9400 Fax: 295-2330  
[www.ultrasci.com](http://www.ultrasci.com)



William J. Leary  
Quality Assurance Manager