



Certificate of Analysis

Pesticide Custom Solution

Product Number: CUS-14103

Page: 1 of 1

Lot Number: CJ-2687

Lot Issue Date: 10-Jul-2012

Expiration Date: 31-Aug-2013

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
kresoxim-methyl	143390-89-0	RM04038	100.3 ± 0.5 µg/mL
trifloxystrobin	141517-21-7	RM03836	100.3 ± 0.5 µg/mL
bifenthrin	082657-04-3	RM00762	100.4 ± 0.5 µg/mL
fludioxonil	131341-86-1	RM05538	100.3 ± 0.5 µg/mL
iprodione	036734-19-7	RM04323	100.4 ± 0.5 µg/mL
lambda-cyhalothrin	091465-08-6	RM05491	100.3 ± 0.5 µg/mL
cypermethrin	052315-07-8	RM01769	100.4 ± 0.5 µg/mL
boscalid	188425-85-6	RM05518	100.3 ± 0.5 µg/mL
fenvalerate	051630-58-1	NT01882	100.4 ± 0.5 µg/mL
deltamethrin	052918-63-5	RM03832	100.3 ± 0.5 µg/mL

Matrix: acetone

Storage: Store at Room Temperature (18-25°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality