



Certificate of Analysis

Custom Standard

Product Number: CUS-14488

Page: 1 of 1

Lot Number: CM-4965

Lot Issue Date: 17-Sep-2015

Expiration Date: 31-Oct-2017

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
dichlobenil	001194-65-6	RM02879	10.0 ± 0.1 µg/mL
diclofop methyl	051338-27-3	RM09811	10.0 ± 0.1 µg/mL
dimethenamid	087674-68-8	RM02938	9.97 ± 0.05 µg/mL
ethofumesate	026225-79-6	RM10744	9.99 ± 0.05 µg/mL
diclobutrazol	075736-33-3	RM02942	9.98 ± 0.05 µg/mL
dichloran	000099-30-9	RM10690	10.0 ± 0.1 µg/mL
etaconazole	060207-93-4	RM09012	9.98 ± 0.05 µg/mL
difenoconazole	119446-68-3	RM10425	9.96 ± 0.05 µg/mL
dialifos	010311-84-9	RM00670	9.99 ± 0.05 µg/mL
dichlofenthion	000097-17-6	RM00208	10.0 ± 0.1 µg/mL
disulfoton	000298-04-4	RM05013	9.96 ± 0.05 µg/mL
ethion	000563-12-2	RM06668	10.0 ± 0.1 µg/mL
ethoprofos	013194-48-4	RM09752	9.98 ± 0.05 µg/mL
etrimfos	038260-54-7	RM10388	10.0 ± 0.1 µg/mL
endosulfan I	000959-98-8	RM00834	9.96 ± 0.05 µg/mL
endosulfan II	033213-65-9	RM07039	10.0 ± 0.1 µg/mL
endosulfan sulfate	001031-07-8	NT00714	10.0 ± 0.1 µg/mL
endrin	000072-20-8	RM00557	9.96 ± 0.05 µg/mL
2,4'-DDE	003424-82-6	RM02765	9.99 ± 0.05 µg/mL
4,4'-DDE	000072-55-9	RM02892	10.0 ± 0.1 µg/mL


Matrix: acetone/ethyl acetate/cyclohexane (50/25/25)

Storage: Store Frozen (-25° to -10°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality