



Certificate of Analysis

Custom Standard

Product Number: CUS-14750

Page: 1 of 1

Lot Number: CK-1576

Lot Issue Date: 13-May-2013

Expiration Date: 31-Dec-2013

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
diallate (total)	002303-16-4	RM06219	50.1 ± 0.3 µg/mL
dichlorvos	000062-73-7	RM06268	50.1 ± 0.3 µg/mL
diclofop methyl	051338-27-3	RM04627	50.1 ± 0.3 µg/mL
dichloran	000099-30-9	RM02869	50.0 ± 0.3 µg/mL
dicofol	000115-32-2	RM03694	50.2 ± 0.3 µg/mL
dieldrin	000060-57-1	RM03957	50.1 ± 0.3 µg/mL
diniconazole	083657-24-3	RM05762	50.0 ± 0.3 µg/mL
dinitramine	029091-05-2	RM02917	50.0 ± 0.3 µg/mL
diphenylamine	000122-39-4	RM02095	50.0 ± 0.3 µg/mL
disulfoton	000298-04-4	RM05013	50.1 ± 0.3 µg/mL
endosulfan I	000959-98-8	RM00834	50.1 ± 0.3 µg/mL
endosulfan II	033213-65-9	RM02905	50.0 ± 0.3 µg/mL
endosulfan sulfate	001031-07-8	NT00714	50.1 ± 0.3 µg/mL
endrin	000072-20-8	RM04132	50.1 ± 0.3 µg/mL
EPTC	000759-94-4	RM02957	50.2 ± 0.3 µg/mL
esfenvalerate (asana)	066230-04-4	RM03731	50.1 ± 0.3 µg/mL
etaconazole	060207-93-4	RM02956	50.0 ± 0.3 µg/mL

Matrix: acetonitrile/acetone

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality