



Certificate of Analysis

Custom Standard

Product Number: CUS-15763

Page: 1 of 1

Lot Number: CL-2774

Lot Issue Date: 17-Jul-2014

Expiration Date: 31-Aug-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
benzene	000071-43-2	RM03830	4020 ± 20 µg/mL
toluene	000108-88-3	RM02555	4018 ± 20 µg/mL
ethylbenzene	000100-41-4	RM00783	4012 ± 20 µg/mL
m-xylene	000108-38-3	RM00053	4013 ± 20 µg/mL
o-xylene	000095-47-6	RM00052	4019 ± 20 µg/mL
p-xylene	000106-42-3	RM02647	4015 ± 20 µg/mL
styrene	000100-42-5	RM06723	4018 ± 20 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4°C.

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality