

Custom Standard

Product Number: CUS-2269

Page: 1 of 1

Lot Number: CF-1768

Lot Issue Date: Apr-2009

Expiration Date: May-2011

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
benzene	000071-43-2	31072	200.6 ± 1.0 µg/mL
ethylbenzene	000100-41-4	01558CC	200.6 ± 1.0 µg/mL
toluene	000108-88-3	45195	200.6 ± 1.0 µg/mL
isopropylbenzene	000098-82-8	EN-00621TG	200.5 ± 1.0 µg/mL
n-propylbenzene	000103-65-1	LO-14503MR	200.6 ± 1.0 µg/mL
4-isopropyltoluene	000099-87-6	PP-05104CP	200.4 ± 1.0 µg/mL
1,2,4-trimethylbenzene	000095-63-6	BO-13528BI	200.5 ± 1.0 µg/mL
1,3,5-trimethylbenzene	000108-67-8	KM-02011HM	200.6 ± 1.0 µg/mL
n-butylbenzene	000104-51-8	AA-28519CO	200.5 ± 1.0 µg/mL
sec-butylbenzene	000135-98-8	MR-11305DN	200.6 ± 1.0 µg/mL
naphthalene	000091-20-3	14205KB	200.8 ± 1.0 µg/mL
MTBE	001634-04-4	38064	200.6 ± 1.0 µg/mL
tert-butylbenzene	000098-06-6	MQ-04010MQ	200.3 ± 1.0 µg/mL
o-xylene	000095-47-6	DO-06834CO	200.6 ± 1.0 µg/mL
m-xylene	000108-38-3	DI-00459CI	200.6 ± 1.0 µg/mL
p-xylene	000106-42-3	03747LN	200.6 ± 1.0 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com



William J. Leary
Quality Assurance Manager