

Custom Standard

Product Number: CUS-8290

Page: 1 of 1

Lot Number: CE-3148

Lot Issue Date: Sep-2008

Expiration Date: Oct-2010

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
benzoic acid	000065-85-0	JM-00103JM	2005 ± 10 µg/mL
4-chloro-3-methylphenol	000059-50-7	MB-061657	2005 ± 10 µg/mL
2-chlorophenol	000095-57-8	CS-09106BQ	2005 ± 10 µg/mL
2,4-dichlorophenol	000120-83-2	HZ-05729EZ	2005 ± 10 µg/mL
2,6-dichlorophenol	000087-65-0	MS-03518LN	2005 ± 10 µg/mL
2,4-dimethylphenol	000105-67-9	BG-14822JF	2005 ± 10 µg/mL
2-methyl-4,6-dinitrophenol	000534-52-1	42136/4	2004 ± 10 µg/mL
2,4-dinitrophenol	000051-28-5	CEE-III-130	2005 ± 10 µg/mL
2-methylphenol	000095-48-7	09416AA	2005 ± 10 µg/mL
3-methylphenol	000108-39-4	BG-03517AG	2005 ± 10 µg/mL
2-nitrophenol	000088-75-5	N980025	2005 ± 10 µg/mL
4-nitrophenol	000100-02-7	HE-5518HE	2005 ± 10 µg/mL
pentachlorophenol	000087-86-5	06324ED	2005 ± 10 µg/mL
phenol	000108-95-2	MY-05714AY	2005 ± 10 µg/mL
2,4,5-trichlorophenol	000095-95-4	DO-07521CO	2005 ± 10 µg/mL
2,4,6-trichlorophenol	000088-06-2	07509ME	2005 ± 10 µg/mL
2,3,4,6-tetrachlorophenol	000058-90-2	ER100206-01	2005 ± 10 µg/mL
4-methylphenol	000106-44-5	TI-09410PI	2005 ± 10 µg/mL

Matrix: methylene chloride (dichloromethane)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com



William J. Leafy
Quality Assurance Manager