



Certificate of Analysis

Pesticides Matrix Spiking Solution

Product Number: CLP-200N

Page: 1 of 1

Lot Number: CM-3696

Lot Issue Date: 13-Jul-2015

Expiration Date: 31-Aug-2017

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
gamma-BHC	000058-89-9	RM02694	2006 ± 10 µg/mL
heptachlor	000076-44-8	RM03111	2004 ± 10 µg/mL
aldrin	000309-00-2	RM00635	2004 ± 10 µg/mL
dieldrin	000060-57-1	RM03957	5010 ± 25 µg/mL
endrin	000072-20-8	RM00557	5011 ± 25 µg/mL
4,4'-DDT	000050-29-3	RM06888	5011 ± 25 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality