



Certificate of Analysis

Pesticides Individual Standard Mixture B

Product Number: CLP-226E

Page: 1 of 1

Lot Number: CM-3562

Lot Issue Date: 10-Jul-2015

Expiration Date: 31-Aug-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	80.0 ± 0.4 ng/mL
beta-BHC	000319-85-7	RM06986	80.1 ± 0.4 ng/mL
delta-BHC	000319-86-8	RM05696	80.2 ± 0.4 ng/mL
cis-chlordane	005103-71-9	RM02742	80.0 ± 0.4 ng/mL
trans-chlordane	005103-74-2	RM02726	80.0 ± 0.4 ng/mL
4,4'-DDE	000072-55-9	RM02891	160.0 ± 0.8 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM00348	160.0 ± 0.8 ng/mL
endosulfan sulfate	001031-07-8	RM06741	160.6 ± 0.8 ng/mL
endrin aldehyde	007421-93-4	RM04131	160.2 ± 0.8 ng/mL
endrin ketone	053494-70-5	RM07329	160.3 ± 0.8 ng/mL
endosulfan II	033213-65-9	RM07039	160.0 ± 0.8 ng/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	80.0 ± 0.4 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM03890	80.2 ± 0.4 ng/mL

Matrix: hexane

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality