



Certificate of Analysis

Pesticides Resolution Check Mixture

Product Number: CLP-242

Page: 1 of 1

Lot Number: CP-1577

Lot Issue Date: 05-Apr-2016

Expiration Date: 31-May-2018

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
trans-chlordane	005103-74-2	RM02726	9.97 ± 0.05 ng/mL
4,4'-DDE	000072-55-9	RM02892	20.1 ± 0.1 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM01256	20.0 ± 0.1 ng/mL
dieldrin	000060-57-1	RM03957	20.0 ± 0.1 ng/mL
endosulfan I	000959-98-8	RM00834	10.0 ± 0.1 ng/mL
endosulfan sulfate	001031-07-8	RM06741	20.1 ± 0.1 ng/mL
endrin ketone	053494-70-5	RM00212	20.1 ± 0.1 ng/mL
methoxychlor	000072-43-5	RM10269	99.8 ± 0.5 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM04082	20.0 ± 0.1 ng/mL


Matrix: hexane

Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality