



Certificate of Analysis

Pesticides Performance Evaluation Mixture

Product Number: CLP-252

Page: 1 of 1

Lot Number: CP-3666

Lot Issue Date: 01-Aug-2016

Expiration Date: 31-Aug-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alpha-BHC	000319-84-6	RM10865	10.0 ± 0.1 ng/mL
beta-BHC	000319-85-7	RM06986	10.0 ± 0.1 ng/mL
gamma-BHC	000058-89-9	RM02694	10.0 ± 0.1 ng/mL
4,4'-DDT	000050-29-3	RM06888	100.3 ± 0.5 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM01256	20.1 ± 0.1 ng/mL
endrin	000072-20-8	RM00557	50.2 ± 0.3 ng/mL
methoxychlor	000072-43-5	RM10269	250.3 ± 1.3 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM09245	20.1 ± 0.1 ng/mL

Matrix: hexane

Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Vincent Wang, Ph.D.
Director of QA/RA

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality