



# Certificate of Analysis

## F004 Solvent List Mixture

**Product Number:** FLM-004

**Page:** 1 of 1

**Lot Number:** CG-0532A

**Lot Issue Date:** 27-Mar-2012

**Expiration Date:** 30-Apr-2015

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
o-cresol	000095-48-7	09416AA	2010 ± 10 µg/mL
m-cresol	000108-39-4	BG03517AG	2010 ± 10 µg/mL
p-cresol	000106-44-5	TI09410PI	2010 ± 10 µg/mL

**Matrix:** methanol (methyl alcohol)

**Storage:** Store at Room Temperature (18-25° C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851-01

ISO 9001:2008  
Registered  
TUV USA, Inc.  
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA  
401-294-9400 Fax: 295-2330  
www.ultrasci.com

---

William J. Leary  
Quality Assurance Manager