




CERTIFICATE OF ANALYSIS

PRODUCT NAME: SIGNAL™ 2-AB LABELING KIT
PRODUCT CODE: GKK-404
LOT NUMBER: 132 237
STORAGE: Room temperature in the dark (DO NOT FREEZE)
EXPIRATION: September 2018


QUALITY CONTROL:

2-AB test labeling: Passed (Specification: "Passed"¹)

Chromatograms of test-labeled and reference oligosaccharide:



Test-labeled Oligosaccharide



Reference Labeled Oligosaccharide

Authorized Signature

1. An oligosaccharide was labeled using a randomly selected kit and the elution profile was compared to a profile of a labeled reference sample of the same oligosaccharide. The retention time and peak fluorescence of the test sample match the reference standard.