

## CERTIFICATE OF ANALYSIS

PRODUCT NAME: SIGNAL™ 2-AB LABELING KIT

PRODUCT CODE: GKK-404

LOT NUMBER: 132 288

STORAGE: Store at Room Temperature (15-30°C) in the dark. DO NOT FREEZE

EXPIRATION: February 2021; the components of this module are usable at least through this month.

COMPONENTS:

Component Name	Quantity/Kit	Lot Number	Expiry
WS0024 Vial A (2-AB Dye)	2 each	W190108b	May 2021
WS0025 Vial B (DMSO)	2 each	W190031	Feb 2021
WS0026 Vial C (Acetic Acid)	2 each	W190083	Apr 2021
WS0027 Vial D (Reductant)	2 each	W190111	May 2021

QC RELEASE:

The components conform to specifications<sup>1</sup>.

---

Authorized Signature

<sup>1</sup> An oligosaccharide was labeled using a randomly selected kit and the elution profile was compared to a profile of a labeled reference sample of the same oligosaccharide. The retention time and peak fluorescence of the test sample match the reference standard.