



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLY-Q ALIGNMENT STANDARDS SET  
PRODUCT CODE: GKSQ-505  
LOT NUMBER: DP17H2103  
STORAGE: Store at -20°C, in the dark. The glycans may be stored on the Gly-Q instrument at room temperature for up to 18 hours without significant impact on performance (cap securely and return to -20°C for longer term storage).  
PACK SIZE: 2 x 100 µl (qualitative standards for glycan identification)  
EXPIRATION: January 2019; the components of this module are usable at least through this month.

### COMPONENTS

Component Code & Name	Quantity/Kit	Lot No.	Expiry
GKSQ-500 Gly-Q Migration Standards (100 µl)	1 each	DP17E2301a	Jan 2019
GKSQ-503 Gly-Q GU Ladder (100 µl)	1 each	DP17E2302a	Jan 2019

See individual component Certificates of Analysis for QC data and Directions for Use.

\_\_\_\_\_  
Authorized Signature