



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP® CLEANUP MODULE (24-ct)  
PRODUCT CODE: GS24-CU  
LOT NUMBER: 944 031  
STORAGE: Store at 2 - 30°C.  
PACK SIZE: 2 x 12 cartridges  
EXPIRATION: July 2020 ; the components of this module are usable at least through this month.

### COMPONENTS

Part No:	Description	Quantity	Lot	Expiry
WS0293	Cleanup (CU) Cartridges	2 x 12 ea	W170104	Jul 2020

### QC RELEASE

The component conforms to specifications<sup>1</sup>.

---

Authorized Signature

- 1 Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.